



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/522,933

01/24/2005

Peter Tass

23158

8318

535 7590 04/13/2009

K.F. ROSS P.C.

5683 RIVERDALE AVENUE

SUITE 203 BOX 900

BRONX, NY 10471-0900

EXAMINER

D ABREU, MICHAEL JOSEPH

ART UNIT

PAPER NUMBER

3762

MAIL DATE

DELIVERY MODE

04/13/2009

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/522,933	TASS, PETER	
	<b>Examiner</b>	<b>Art Unit</b>	
	Michael D'Abreu	3762	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 31 January 2009.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-38 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-38 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)          | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)          | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____  | 6) <input type="checkbox"/> Other: _____                          |

### **DETAILED ACTION**

1. This Office action is responsive to communications filed on 31 January 2009.

#### ***Response to Amendment***

2. Applicant's request for reconsideration of the finality of the rejection of the last Office action is persuasive and, therefore, the finality of that action is withdrawn. This Office action has been made final since it is based on amended claims. The claims were amended due to the first Office action.

#### ***Claim Rejections - 35 USC § 103***

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. Claims 1-7, 9-19, 24-33, 37, & 38 are rejected under 35 U.S.C. 103(a) as being unpatentable over John (3780724).
5. Regarding Claims 1-6 & 38, John discloses a device capable of testing a patient for neuronal rhythmic activity comprising of a control unit inherently evoking physiological brain activity (e.g. Column 3, lines 15-28; Claim 1). The device has several testing methods (e.g. Col 7-8), including one which generates a predictable periodic succession of pulses followed by another periodic succession of pulses (e.g. Col 9, lines 3-21) with or without a pre-determined rest or time delay (e.g. Col 8, lines 1-9). John

Art Unit: 3762

specifies that the patient is stimulated visually and acoustically via a light or speaker as a stimulator (e.g. Col 3, lines 29-34). Means for detecting brain activity are further disclosed through scalp EEG electrodes (e.g. Col 3, lines 15-20), which are connected to the control unit by an amplifier (e.g. Col 3, lines 20-27). Furthermore, the device has feedback of patient reaction and connected to the control unit (e.g. Col 2, lines 21-47).

6. John's device generates a periodic succession of pulses for testing memory; however, it is not indicated that these pulses can control and abruptly desynchronize the phase dynamic of neuronal rhythmic activity. The examiner notes that short term memory itself is a neuronal activity and visual stimulations which are repeated to a patient create neuronal activity simply based on the rhythm of the stimuli. Furthermore, the applicant uses the similar types of visual stimuli as John to control neuronal rhythmic activity. Therefore, it would have been obvious to one of ordinary skill to modify the device of John to direct the pulses towards neuronal rhythmic activity to yield the predictable results of providing stimuli which can control the phase dynamic of neuronal rhythmic activity.

7. Concerning Claims 7, 8, & 10, John's device carries out frequency scans (e.g. Col 4, lines 15-20), quantifies neuronal activity (e.g. Col 2, lines 23-30), and is designed to have the stimulator directly connected and activated by the control unit (e.g. Col 3, lines 62-67).

8. With respect to Claims 11-13, the apparatus as disclosed by John describes a T test computer, capable of investigating signals measured by the sensors through the use of the wavelet analysis (e.g. Col 4, lines 15-20). Furthermore, the device registers

Art Unit: 3762

the change in the amplitude of the rhythm to be excited by recording that response (e.g. Col 4, lines 23-25).

9. Claims 14 & 15, John's device has various testing methods including means for carrying out an entrainment (e.g. Col 7, lines 5-20) and desynchronization (e.g. Col 7, lines 44-50).

10. Regarding Claims 16-19, the T test computer in John's device is configured to test the quality of the entrainment (e.g. Col 4, lines 16-40) by determining the phase of the neuronal rhythms by matching the signals. The phase and amplitude of the neuronal activity is then evaluated (e.g. Col 4, 41-62).

11. With respect to Claims 24-27 and 33, John's device determines the vulnerable phase of the neural signal as defined by the applicant by varying the time spacing between the last excitation of the entrainment and the desynchronizing excitation signal (e.g. Col 8, lines 10-24). In one method, the device also changes the variation in time spacing for different values of intensity and in another method, increases the intensity in equidistant steps (e.g. Col 8, lines 18-22).

12. Regarding Claims 28 and 29, the prior art enables optimum stimulation parameters to be determined through the comparison of the results of multiple testing stimulations from where a minimization of the amplitude can be obtained (e.g. Cols 7-9).

13. Claim 37 is anticipated by John's device as the stimulation is monitored by a program, in conjunction with a timer and switch (e.g. Col 3, 34-40).

Art Unit: 3762

14. Claims 9 and 30-32 are rejected under 35 U.S.C. 103(a) as being unpatentable over John (3780724) in view of Nagano (5771443). John discloses the device substantially as claimed; however, specific means of quantifying neuronal activity or detecting stimulation parameters are not disclosed. The examiner notes that the applicant fails to disclose a clear or definitive advantage behind using the Hilbert transformation, matching a sine function, or integrating amplitude of power spectrum over frequency band over any other form of wavelet analysis. Nagano discloses the use of Hilbert transformation and other forms of wavelet analysis in the measurement and analysis of frequency deviation (e.g. Col 7). It would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the method taught by John, in view of the wavelet analysis of Nagano, to provide the predictable results of a more detailed examination of the neural activity.

15. In reference to Claim 23, John teaches the method as substantially as claimed. Although the amplitude response is analyzed by John's device in order to find marked differences between the responses (e.g. Col 7, lines 30-43), John does not plot the data as amplitude resetting curves in the analysis of neuronal activity. It would have been obvious to one of ordinary skill in the art, at the time the invention was made, to plot these data points in order to provide the predictable results of visually presenting the responses as figures rather than numbers. The applicant makes a graph of the amplitude data points - but the computer is still analyzing the data points in the same manner as the prior art. Plotting a set of data points on a graph to visually present a curve is common in basic experimentation and is not any sort of improvement.

Art Unit: 3762

16. Claims 20-22 and 34-36 are rejected under 35 U.S.C. 103(a) as being unpatentable over John (3780724) in view of Czeisler (5545192). John discloses the device substantially as claimed. However, John's device does not employ the use of phase resetting curves in neural analysis. Czeisler uses phase resetting curves in the analysis of circadian rhythm. It would have been obvious to one of ordinary skill in the art to modify the method taught by John, to include phase resetting curves in order to provide the predictable results of projecting and analyzing the neural activity of the patient. Furthermore, John's device discloses the use of a T-test computer which is able to quantitatively analyze the response and phase dynamics of the desynchronizing neuronal activity both before and after stimulation (e.g. Cols 4-6).

### ***Response to Arguments***

Applicant's arguments filed 31 January 2009 have been fully considered but they are not persuasive. In response to Applicant's argument that the succession of pulses do not control the phase dynamic of the neuronal rhythmic activity, the fact that Applicant uses the visual stimulus for a different purpose does not alter the conclusion that its use in a prior art device would be prima facie obvious from the purpose disclosed in the reference. Short term memory itself is a neuronal activity and it is obvious that the repetition of a large and small square 30-100 times would control the phase dynamic of the neuronal rhythmic activity. Furthermore, by suddenly changing the repeating display to a large and small round figure, the rhythmic activity induced by the previous display would be desynchronized (e.g. Col 9, lines 3-21).

### ***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael D'Abreu whose telephone number is (571)270-3816. The examiner can normally be reached on Monday - Friday, 0600 - 1630 EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on (571) 272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.



Art Unit: 3762

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/M. D./  
Examiner, Art Unit 3762

/George R Evanisko/  
Primary Examiner, Art Unit 3762